~ Calibration Certificate ~

353B33 Model Number: Serial Number: LW184234 ICP® Accelerometer Description: **PCB** Manufacturer: Method: Back-to-Back Comparison AT401-3

Calibration Data

Sensitivity @ 100 Hz

101.6 mV/g

Output Bias

10.4 VDC

 (10.36 mV/m/s^2)

Transverse Sensitivity

0.6 %


Discharge Time Constant

1.0 seconds

Resonant Frequency

25.7 kHz

Sensitivity Plot



Hz

Data Points

Frequency (Hz)	Dev. (%)	Frequency (Hz)	Dev. (%)
10	-0.3	300	0.2
15	-0.4	500	0.3
30	-0.2	1000	0.5
50	-0.1	3000	2.0
REF. FREQ.	0.0	4000	3.2

Mounting Surface: Beryllium w/Silicone Grease Fastener 10-32 Female Fixture Orientation: Vertical Acceleration Level (pk): 10-10 g (pk): 10 m/s?)

The acceleration level may be limited by shaker displacement at low frequencies. If the listed level cannot be obtained, the calibration system uses the following formula to set the vibration amplitude; Acceleration Level (g) = 0-008 x (freq)². The gravitational constant used for calculations by the calibration system is; 1 g = 9.80665 m/s².

Condition of Unit

As Found:

As Left:

n/a

New Unit, In Tolerance

Notes

- 1. Calibration is NIST Traceable thru Project 683/283498 and PTB Traceable thru Project 10065.
- 2. This certificate shall not be reproduced, except in full, without written approval from PCB Piezotronics, Inc.
- 3. Calibration is performed in compliance with ISO 9001, ISO 10012-1, ANSI Z540.3 and ISO 17025.
- 4. See Manufacturer's Specification Sheet for a detailed listing of performance specifications.
- 5. Measurement uncertainty (95% confidence level with coverage factor of 2) for frequency ranges tested during calibration are as follows: 5-9 Hz; +/- 2.0%, 10-99 Hz; +/- 1.5%, 100-1999 Hz; +/- 1.0%, 2-10 kHz; +/- 2.5%.

Date: ____

4/6/2015



VIBRATION DIVISION Headquarters: 3425 Walden Avenue, Depew, NY 14043 Calibration Performed at: 10869 Highway 903, Halifax, NC 27839

TEL: 888-684-0013 - FAX: 716-685-3886

www.pcb.com

CAL76-3511169622.997+0

PAGE I of I